

Mouse/Rat PDGF-BB ELISA Instructions

Cat:EM0062

Content

	CAT	Volume
① CP (Coated Plate)	EM0062CP	96 well
② S (Standard)	EM0062S	2 vial
③ SD (Sample Diluent)	ESD21	15 ml/bottle
④ DA (Detect Antibody)	EM0062DA	6 ml/bottle
⑤ SH (Streptavidin-HRP)	ESH01	12 ml/bottle
⑥ AB (Assay Buffer 1×)	EAB01	12 ml/bottle
⑦ TS (TMB Substrate)	ETS01	12 ml/bottle
⑧ SS (Stop Solution)	ESS01	12 ml/bottle
⑨ WB (Wash Buffer 10×)	EWB01	50 ml/bottle
⑩ SF (Sealer Film)	ESF01	6 pieces

NOTE: After the kit is opened, the stabilization period of each content is 30 days, so please use the kit within 30 days after opening.

REAGENT PREPARATION


Standard Curve Preparation:

The **SD** (Sample Diluent) is used for serum and plasma. Other sample type, prepare the standard curve with whatever buffer (SPB, Sample Prepared Buffer) is used to prepare the sample, such as cell culture supernatant, tissue grinding liquid, cell lysate, etc. Urine sample use **AB** (Assay Buffer) prepare standard curve.

Reconstitute mouse/rat PDGF-BB Standard by addition of distilled water as S. Reconstitution volume is stated on the label of the standard vial. Swirl or mix gently to insure complete and homogeneous solubilization (concentration of reconstituted standard = 100000pg/ml).

Allow the standard to reconstitute for 10-30 minutes. Mix well prior to making dilutions.

The mouse/rat PDGF-BB Standard EM0062S 100000pg/ml 30 μ l + 270 μ l **SD** (Sample Diluent) serves as the high standard (10000pg/ml). Pipette 150 μ l of **SD** (Sample Diluent) into each tube. Use the high standard to produce a 2:3 dilution series. Mix each tube thoroughly before the next transfer. **SD** (Sample



Washing Buffer (1×) Preparation


Pour entire contents (50 ml) of the **Washing Buffer Concentrate** (10×) into a clean 500 ml graduated cylinder. Bring to final volume of 500 ml with glass-distilled or deionized water. Transfer to a clean wash bottle and store at 2 to 25°C.

ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use.

- ① Prepare all reagents and working standards as directed in the previous sections.
- ② Remove excess **CP** (Coated Plate) strips from the plate frame, return them to the foil pouch and reseal.
- ③ Add 50 μ l of **AB** (Assay Buffer) to each well.
- ④ Add 10 μ l of **Standard or sample** per well. Ensure reagent addition is uninterrupted and completed within 15 minutes.
- ⑤ Add 50 μ l of **DA** (Detect Antibody) to each well.
- ⑥ Cover with an **SF** (Sealer Film). Incubate at room temperature (18 to 25°C) for 1 hour on a microplate **shaker** set at 500 rpm.
- ⑦ Aspirate each well and **wash**, repeating the process four times. Wash by filling each well with **WB** (Washing Buffer 300 μ l). Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining **WB** (Washing Buffer) by aspirating or decanting. Invert the plate and **blot** it against clean paper towels.
- ⑧ Add 100 μ l of **SH** (Streptavidin-HRP) to each well.
- ⑨ Cover with a new **SF** (Sealer Film). Incubate at room temperature (18 to 25°C) for 30 min on a microplate **shaker** set at 500 rpm.
- ⑩ Repeat aspiration/**wash** as in step 7.
- ⑪ Add 100 μ l of **TS** (TMB Substrate) to each well. Incubate for 5-30 minutes at room temperature.
- ⑫ Add 100 μ l of **SS** (Stop Solution) to each well.
- ⑬ Determine the optical density within 30 minutes, using microplate **reader** set to 450 nm corrected with 570 nm or 630 nm.

TYPICAL DATA



pg/ml	O.D.	Average	Corrected
0.00	0.0068	0.0066	0.0067
40.96	0.0202	0.0190	0.0129
102.40	0.0481	0.0460	0.0404
256.00	0.1329	0.1363	0.1279
640.00	0.3743	0.3821	0.3715
1600.00	0.9517	0.9508	0.9446
4000.00	2.2200	2.2270	2.2168
10000.00	3.8750	3.9774	3.9195

SENSITIVITY

The minimum detectable dose (MDD) of mouse/rat PDGF-BB is typically less than 5.18 pg/ml (10 µl of sample volume).

The MDD was determined by adding two standard deviations to the mean optical density value of ten zero standard replicates and calculating the corresponding concentration.

PRECISION

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested twenty-two times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Sample Number	Intra-assay Precision			Inter-assay Precision		
	S1	S2	S3	S1	S2	S3
22	22	22	22	6	6	6
Average (pg/ml)	247.5	1029.7	3128.0	246.6	1028.9	3212.6
Standard Deviation	7.3	29.6	116.9	5.9	24.4	85.7
Coefficient of Variation (%)	2.9	2.9	3.7	2.4	2.4	2.7

RECOVERY

The spike recovery was evaluated by spiking 3 levels of mouse/rat PDGF-BB into health mouse/rat serum sample. The un-spiked serum was used as blank in this experiment.

The recovery ranged from 80% to 115% with an overall mean recovery of 87%.

LINEARITY

To assess the linearity of the assay, five samples were spiked with high concentration of PDGF-BB in mouse/rat serum and diluted with Sample Diluent to produce samples with values within the dynamic range of the assay.

The linearity ranged from 82% to 108% with an overall mean recovery of 90%.

SAMPLE VALUES

Serum/Plasma – Thirty samples from apparently healthy mice and thirty samples from apparently healthy rats were evaluated for the presence of mouse/rat PDGF-BB in this assay. No medical histories were available for the donors.

Sample Matrix	Sample Evaluated	Range (pg/ml)	Detectable %	Mean of Detectable (pg/ml)
Mouse Serum	30	1150.55-3477.50	100	1718.04
Rat Serum	30	841.58-7059.93	100	2701.44

n.d. = non-detectable. Samples measured below the sensitivity are considered to be non-detectable.