

Mouse alpha 1 antitrypsin ELISA Instructions

Cat: EMY005

Content

	CAT	Volume
1 CP (Coated Plate)	EMY0052CP	96 well
2 S (Standard)	EMY0052S1	2 vial
3 SD (Sample Diluent)	ESD01	15 ml/bottle
4 DA-H (Detect Antibody-HRP 100×)	EMY0052DA-H	1 vial
5 DD (Detect Antibody Diluent)	EDD02	6 ml/bottle
6 AB (Assay Buffer 1×)	EAB01	12 ml/bottle
7 TS (TMB Substrate)	ETS01	12 ml/bottle
8 SS (Stop Solution)	ESS01	12 ml/bottle
9 WB (Wash Buffer 10×)	EWB01	50 ml/bottle
10 SF (Sealer Film)	ESF01	6 pieces

NOTE: After the kit is opened, the stabilization period of each content is 30 days, so please use the kit within 30 days after opening.

Sample Dilution

Samples such as serum, plasma require at least a 100-fold dilution into Sample Diluent. A suggested 100-fold dilution is 5 μ l of sample + 495 μ l of Sample Diluent.


REAGENT PREPARATION

Standard Curve Preparation:

Reconstitute Mouse alpha 1 antitrypsin Standard by addition of distilled water as S1. Reconstitution volume is stated on the label of the standard vial. Swirl or mix gently to insure complete and homogeneous solubilization (concentration of reconstituted standard = 10000ng/ml).

Allow the standard to reconstitute for 10-30 minutes. Mix well prior to making dilutions.

The Mouse alpha 1 antitrypsin Standard EMY0052S1 as the high standard (10000ng/ml). Pipette 200 μ l of SD into each tube. Use the high standard to produce a 1:2 dilution series. Mix each tube thoroughly before the next transfer. SD serves as the zero standard (0 ng/ml).



Washing Buffer (1×) Preparation

Pour entire contents (50 ml) of the **Washing Buffer Concentrate** (10×) into a clean 500 ml graduated cylinder. Bring to final volume of 500 ml with glass-distilled or deionized water. Transfer to a clean wash bottle and store at 2 to 25°C.

1×DA Preparation:

Mix well prior to making dilutions.


Make a 1:100 dilution of the concentrated Detect Antibody solution with **DD** (Detect Antibody Diluent) in a clean plastic tube as needed according to the Standards and Samples.

ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use.

- 1 Prepare all reagents and working standards as directed in the previous sections.
- 2 Remove excess **CP** (Coated Plate) strips from the plate frame, return them to the foil pouch and reseal.
- 3 Add 50 μ l of **AB** (Assay Buffer) to each well.
- 4 Add 10 μ l of **Standard or sample** per well. Ensure reagent addition is uninterrupted and completed within 15 minutes.
- 5 Add 50 μ l of **DA-H** (Detect Antibody-HRP) to each well.
- 6 Cover with an **SF** (Sealer Film). Incubate at room temperature (18 to 25°C) for 30 min on a microplate **shaker** set at 500 rpm.
- 7 Aspirate each well and **wash**, repeating the process four times. Wash by filling each well with **WB** (Washing Buffer 300 μ l). Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining **WB** (Washing Buffer) by aspirating or decanting. Invert the plate and **blot** it against clean paper towels.
- 8 Add 100 μ l of **TS** (TMB Substrate) to each well. Incubate for 5-30 minutes at room temperature.
- 9 Add 100 μ l of **SS** (Stop Solution) to each well.
- 10 Determine the optical density within 30 minutes, using microplate **reader** set to 450 nm corrected with 570 nm or 630 nm.

TYPICAL DATA



ng/ml	O.D.	Average	Corrected
0.00	0.0085	0.0087	0.0086
13.72	0.0168	0.0182	0.0175
41.15	0.0345	0.0348	0.0347
123.46	0.0903	0.0873	0.0888
370.37	0.2459	0.2427	0.2443
1111.11	0.6791	0.6760	0.6776
3333.33	1.7800	1.7870	1.7749
10000.00	3.7750	3.8150	3.7950
			3.7864

SENSITIVITY

The minimum detectable dose (MDD) of Mouse alpha 1 antitrypsin is typically less than 1.57 ng/ml .

The MDD was determined by adding two standard deviations to the mean optical density value of ten zero standard replicates and calculating the corresponding concentration.

PRECISION

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Sample Number	Intra-assay Precision			Inter-assay Precision		
	S1	S2	S3	S1	S2	S3
Average (ng/ml)	203.5	1135.4	4091.5	165.0	911.8	3156.6
Standard Deviation	15.0	70.0	273.9	8.7	57.5	117.3
Coefficient of Variation (%)	7.4	6.2	6.7	5.3	6.3	3.7

RECOVERY

The spike recovery was evaluated by spiking 3 levels of Mouse alpha 1 antitrypsin into health human serum sample. The un-spiked serum was used as blank in this experiment.

The recovery ranged from 90% to 120% with an overall mean recovery of 102%.

LINEARITY

To assess the linearity of the assay, five samples were spiked with high concentration of Neupilin-1/NRP1 in human serum and diluted with Sample Diluent to produce samples with values within the dynamic range of the assay.

The linearity ranged from 92% to 118% with an overall mean recovery of 106%.

SAMPLE VALUES

Serum/Plasma – Thirty samples from apparently healthy volunteers were evaluated for the presence of Mouse alpha 1 antitrypsin in this assay. No medical histories were available for the donors.

Sample Matrix	Sample Evaluated	Range (ug/ml)	Detectable %	Mean of Detectable (ug/ml)
Serum	30	210.39-335.68	100	288.48

n.d. = non-detectable. Samples measured below the sensitivity are considered to be non-detectable.